J Nanomed Nanotechnol 2018, Volume 9 DOI: 10.4172/2157-7439-C5-080

conferenceseries.com

20th Asia-Pacific Nanotechnology Congress

July 23-24, 2018 Sydney, Australia

Confined nanoscale geometries to enhance sensitivity of plasmonic immunoassays

Rishabh Rastogi

Luxembourg Institute of Science and Technology, Luxembourg

Sensitive transduction of bio-molecular binding events on chip carries profound implications to the outcome of a range of biological sensors. This includes biosensors that address both research as well as diagnostic questions of clinical relevance like profiling of biomarkers, protein expression analysis, drug or toxicity screening and drug-efficacy monitoring. Nano structured biosensors constitute a promising advance in this direction owing to their ability in catering to better sensitivity, response times and miniaturization. Plasmonic sensors are particularly interesting among nano-biosensors as they exploit light matter interactions in the nanoscale to transduce bio-recognition events with high sensitivity and miniaturized measurement footprints. Examples of plasmonic sensors include Localized Surface Plasmon Resonance Spectroscopy (LSPR), Surface Enhanced Raman Spectroscopy (SERS) and Metal-Enhanced Fluorescence (MEF). The performance of the plasmonic sensors critically relies on ability to engineer nanoscale geometric attributes at length scales that typically overlap with the size of small proteins. Such geometries invariably introduce constraints on the molecular binding response, thus altering the interaction outcomes, viz. density and kinetics of adsorption, molecular orientations, in a manner that would impact the resulting optical response. A careful engineering of the nanoscale geometries can simultaneously take advantage of EM field enhancements together with molecular interaction within nanoscale geometries. The project aims at an engineered nanoscale interface with geometry tailored to simultaneously favour molecular adsorption and plasmonic enhancements for application to plasmonic sensors based on surface-enhanced Raman and fluorescence spectroscopies.

rishabh.rastogi@list.lu